



Figure 1. Glucose meter validation study conducted at St Elizabeth Medical Center, September 2013 through January 2014. Coefficient of correlation = 0.989; slope = 0.975; y intercept = 1.256; 99% confidence interval 11; n = 350.