

Potential Risk	Control Process and Risk Mitigation Strategies
Instrument failure	System and result memory integrity checks with every test.
Instrument maintenance	Testing demonstrates results after cleaning and disinfection cycles simulate 3 years.
Instrument reliability	Linearities performed every 6 months; subset correlated with chemistry analyzer. Measurement range 10–600 mg/dL.

Table 4. Risk assessment for measuring system.